

Mono-planar T-Hex: Speed and flexibility for high-resolution 3D imaging

Journal Article**Author(s):**

Engel, Maria; [Kasper, Lars](#) ; Wilm, Bertram; Dietrich, Benjamin; Patzig, Franz; Vionnet, Laetitia; Pruessmann, Klaas P.

Publication date:

2021-01

Permanent link:

<https://doi.org/10.3929/ethz-b-000502401>

Rights / license:

[Creative Commons Attribution 4.0 International](#)

Originally published in:

Magnetic Resonance in Medicine 87(1), <https://doi.org/10.1002/mrm.28979>

Mono-planar T-Hex: Speed and flexibility for high-resolution 3D imaging

Maria Engel¹  | Lars Kasper^{1,2}  | Bertram Wilm¹  | Benjamin Dietrich¹ | Franz Patzig¹  | Laetitia Vionnet¹ | Klaas P. Pruessmann¹ 

¹Institute for Biomedical Engineering, ETH Zurich and University of Zurich, Zurich, Switzerland

²Translational Neuromodeling Unit, IBT, University of Zurich and ETH Zurich, Zurich, Switzerland

Correspondence

Maria Engel, Institute for Biomedical Engineering, ETH Zurich and University of Zurich, ETZ F 87, Gloriastrasse 35, 8092 Zurich, Switzerland.
Email: engel@biomed.ee.ethz.ch

Purpose: The aim of this work is the reconciliation of high spatial and temporal resolution for MRI. For this purpose, a novel sampling strategy for 3D encoding is proposed, which provides flexible k-space segmentation along with uniform sampling density and benign filtering effects related to signal decay.

Methods: For time-critical MRI applications such as functional MRI (fMRI), 3D k-space is usually sampled by stacking together 2D trajectories such as echo planar imaging (EPI) or spiral readouts, where each shot covers one k-space plane. For very high temporal and medium to low spatial resolution, tilted hexagonal sampling (T-Hex) was recently proposed, which allows the acquisition of a larger k-space volume per excitation than can be covered with a planar readout. Here, T-Hex is described in a modified version where it instead acquires a smaller k-space volume per shot for use with medium temporal and high spatial resolution.

Results: Mono-planar T-Hex sampling provides flexibility in the choice of speed, signal-to-noise ratio (SNR), and contrast for rapid MRI acquisitions. For use with a conventional gradient system, it offers the greatest benefit in a regime of high in-plane resolution <1 mm. The sampling scheme is combined with spirals for high sampling speed as well as with more conventional EPI trajectories.

Conclusion: Mono-planar T-Hex sampling combines fast 3D encoding with SNR efficiency and favorable depiction characteristics regarding noise amplification and filtering effects from T_2^* decay, thereby providing flexibility in the choice of imaging parameters. It is attractive both for high-resolution time series such as fMRI and for applications that require rapid anatomical imaging.

KEYWORDS

3D encoding, algebraic image reconstruction, magnetic field monitoring, spiral imaging

This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2021 The Authors. *Magnetic Resonance in Medicine* published by Wiley Periodicals LLC on behalf of International Society for Magnetic Resonance in Medicine

1 | INTRODUCTION

High-resolution volume imaging demands high time efficiency of encoding and acquisition, particularly when temporal resolution is also critical, as in fMRI¹⁻³ and fQSM,^{4,5} or when contrast preparation is time-consuming as, for example, in arterial spin labeling (ASL).⁶ For high speed of volume imaging, 3D acquisition is well suited, benefiting from array encoding in three dimensions. One common strategy for fast 3D acquisition is to stack EPI or spiral readouts in 3D k-space, covering one entire k-space plane at a time.⁷ This is possible up to a resolution limit, typically somewhat below 1 mm, set by signal decay, intra-voxel dephasing and gradient performance.

For higher resolution with k-space stacks of planar readouts, the conventional option is to cover each plane by multiple, interleaved shots.⁸ However, integer increments in the number of interleaves cause abrupt sequence changes that may be disadvantageous. For instance, moving from one to two shots halves the readout time, greatly reducing TR and thus causing leaps in optimal flip angle, contrast, and SNR efficiency. Actual optimization of contrast and SNR yield will generally favor a certain TR regardless of the targeted resolution. Furthermore, due to sequence overhead such as preparation modules, RF excitation, and spoiling, integer steps in the number of interleaves also come with substantial increases in total scan time.

High-resolution 3D imaging thus calls for covering each k-space plane with non-integer numbers of shots larger than one. This problem bears resemblance to the reverse challenge of sampling more than one k-space plane per shot in the medium- to low-resolution regime. The latter case has recently been tackled by multi-planar spirals and EPI readouts along tilted hexagonal grids in the indirect dimensions (T-Hex).⁹

In this work, we propose a modified form of the T-Hex concept that addresses the high-resolution scenario, achieving the desired fractional coverage per shot with planar readouts. We demonstrate that mono-planar T-Hex increases the flexibility of sequence timing with EPI and spiral acquisition. With spirals, it achieves whole-brain T_2^* -weighted imaging with sub- μ l resolution in less than 5 s, using a regular gradient system.

2 | METHODS

2.1 | Trajectory design

2.1.1 | Tilted hexagonal grids

Fast 3D Fourier encoding is often achieved by stacking identical 2D readouts in the third dimension. EPI and Archimedean

spirals then form 2D Cartesian grids in planes orthogonal to the read direction. These planes, which are parallel for EPI and radial about the principal axis of the stack for spirals, shall be referred to as phase-encoding (PE) planes. However, Cartesian sampling is usually suboptimal. Structures with approximately circular or elliptical support are sampled more efficiently on hexagonal grids.¹⁰⁻¹³ In parallel imaging with undersampling, the greater isotropy of hexagonal patterns improves the conditioning of image reconstruction for a given mean sampling density.¹⁴⁻¹⁷

Spiral and EPI trajectories can readily be stacked in such a way that they cover hexagonal grids in their PE planes (Figure 1A). However, due to the finite feasible length of individual readouts, one shot may not suffice to cover an entire nominal k-space plane when targeting high resolution. The common solution to this problem is to segment each plane into N interleaved shots. However, due to the inflexibility in time allocation arising from N being an integer, it is often impossible to optimize the sequence timing for contrast and SNR as discussed more closely in the Supporting Information.

Towards greater flexibility of segmentation, a key parameter is the PE increment d of each individual shot, which is the spacing of readout lines in EPI and the radial increment per turn for a spiral readout. For conventional segmentation, d is constrained to integer multiples of the lattice constant Δk as shown in Figure 1A. However, other distances occur on the hexagonal grid as illustrated in Figure 1B. Defining the hexagonal grid by the basis vectors \vec{e}_1 and \vec{e}_2 , a greater variety of vector lengths results from forming integer combinations of \vec{e}_1 and \vec{e}_2 :

$$\vec{v} = M_1 \cdot \vec{e}_1 + M_2 \cdot \vec{e}_2 \quad (1)$$

Due to the symmetries of the grid, all possible distances are represented by vectors from the origin to points in the first dodecant marked yellow in Figure 1B ($M_1 \geq M_2 \geq 0$).

This greater choice of distances can be exploited to increase the flexibility of segmentation by selecting a grid point with favorable distance d from the origin (eg, $M_1 = 2$, $M_2 = 1$, marked red in Figure 1B) and tilting the grid such that the corresponding vector \vec{v} is orthogonal to the principal axis of the stack (2nd PE axis in Figure 1C). For all choices of \vec{v} , the points on the tilted hexagonal grid fall onto a finer, rectilinear grid (green lines in Figure 1D), as derived in the Appendix of Ref. [9].

For optimal imaging performance, the eventual sampling density must account for the required FOV and the parallel imaging capability of the array used. Either or both may be anisotropic, calling for different densities in the two PE directions. To implement these, the tilted hexagonal pattern is rescaled in the two dimensions and cropped according to the desired resolution.




FIGURE 1 A, Gray dots depict part of the hexagonal grid in a transverse section of a stack of spirals or echo-planar readouts. Blue dots mark the revolutions acquired within a shot for $N = 1, 2,$ and 3 shots per k -space plane. B, The hexagonal grid is described by oblique coordinates. All existing distances between grid points are represented in distances of the origin to points in one dodecant (yellow wedge), including its edges. Blue rings mark the distances reflected in (A), and the red ring highlights a distance not available with integer N . C, This distance is used when tilting the grid with respect to the rotational axis of the stack. The FE lines or spiral revolutions acquired in three subsequent shots are highlighted in red, orange, and yellow. D, The hexagonal grid is tilted as shown in (C). Red and yellow dots mark the points of the hexagonal grid that are covered by two subsequent shots. Green lines in the background show that all points of the hexagonal grid lie on nodes of a finer, rectilinear lattice, from which the starting angle of each spiral shot can be derived. Subsequent shots start with linearly progressing angles. The pattern is repetitive. E, Background: An entire mono-planar T-Hex stack of spirals cut open. The uppermost shot and the sites of puncture of the spirals are marked in dark gray. The resulting hexagonal grid shows the tilt as visualized in (C) and (D). In a representative patch, grid points that belong to spiral shots with the same phasing are colored the same. Foreground: Cross-section corresponding to the stack in the upper panel, the acquisition time point is color-coded, indicating a smooth T_2^* filter. Owing to the cross-sectional depiction, “holes” appear on the central k -space axis. In fact, this region is equally uniformly sampled, since all spirals start from the rotational axis of the cylinder

2.1.2 | Application to spirals and EPI

The Archimedean spiral underlying a stack of spirals (Figure 1E) needs to be designed only once. It is then rotated and shifted along the 2nd PE direction according

to the respective T-Hex pattern, involving fixed shot-to-shot phase and shift increments. Like in the case of multi-planar T-Hex readouts,⁹ identical radial progression of all interleaves ensures smooth T_2^* weighting. Gained flexibility of the increment d in the 1st PE direction corresponds

to flexibility in slab thickness in the multi-planar case. Similarly, the repetitiveness of blipping in the multi-planar case corresponds to repetition of the sequence of phase offsets in the mono-planar scenario. The number of shots within this sequence reads

$$\frac{2(M_1^2 + M_2^2 + M_1M_2)}{\gcd(2M_2 + M_1, 2M_1 + M_2)}$$

In the example shown in Figure 1D, for instance, the sequence has length 14. Blipping as such and any related deceleration are not necessary in the mono-planar case.

All of these considerations apply equally to the EPI variant, with shot-to-shot in-plane shifts instead of rotations as well as corresponding echo time shifts for smooth T_2^* weighting. The echo time shift is proportional to the ratio of the in-plane shift (c in Figure 1D) and the spacing d and scales with the temporal spacing of readout lines, analogous to conventional segmented EPI.¹⁸

Based on the resulting k-space trajectories, the gradient waveforms were computed so as to minimize their duration within gradient-strength and slew-rate constraints.¹⁹

2.2 | Speed and flexibility

To assess the utility of the proposed approach, resulting scan times and underlying acquisition times per shot, T_{AQ} , were calculated for a typical whole-brain imaging scenario with FOV = $24 \times 24 \times 12$ cm³ and varying in-plane resolution of 0.4 mm, 0.6 mm, and 0.8 mm, and 2 mm through-plane resolution. Stacks of mono-planar T-Hex spirals were generated as described above, varying the tilt according to all points in the yellow wedge in Figure 1B with d up to $7.6 \Delta k$, which corresponds to $M_1 = 7$, $M_2 = 1$. For comparison, spirals were also stacked conventionally, forming Cartesian grids in PE planes. Hexagonal and Cartesian grids were chosen such as to encode the same ellipse in the image domain with net undersampling by a factor of $R = 8$. An overhead of 13 ms per shot was included to account for typical excitation, fat suppression, and spoiling modules.

TABLE 1 Sequence parameters

Trajectory type	FOV [cm ³]	R	Resolution	TR [ms]	TE [ms]	No. of shots	Total scan duration
T-Hex spiral-out	$24 \times 24 \times 12$	8	$0.6 \times 0.6 \times 2$ mm ³	56	15	86	4.8 s
T-Hex spiral-out	$24 \times 24 \times 12$	8	$0.6 \times 0.6 \times 2$ mm ³	66	25	86	5.7 s
T-Hex EPI	$24 \times 24 \times 12$	8	$0.7 \times 0.7 \times 2$ mm ³	50	20	86	4.3 s
Cartesian spin-warp	$24 \times 24 \times 13.2$	1	1.5 mm ³	31	1.95-6.80	10500	5 min 26 s

2.3 | In-vivo experiments

Experiments were carried out on a 7T whole-body MRI system (Achieva, Philips Healthcare, Best, The Netherlands), using a quadrature-transmit and 32-channel receive head setup (Nova Medical, Wilmington, USA). The system was operated in a mode offering gradient amplitude up to 31 mT/m and a maximum slew-rate of 200 T/m/s.

Mono-planar spiral and EPI T-Hex readouts were included in a common gradient echo sequence (parameters shown in Table 1, $M_1 = 2$, $M_2 = 1$). Additional 3D multi-echo spin-warp scans served for B_0 and coil sensitivity mapping. T-Hex EPI included a SPIR fat suppression module.²⁰ K-space trajectories were pre-recorded using a field camera (Skope Magnetic Resonance Technology, Zurich, Switzerland) made from 16 ¹H NMR field probes ($T_1 = 62$ ms, $T_2^* = 36$ ms),²¹⁻²⁴ operated with the console described in Ref. [25].

Data were collected from two healthy volunteers according to the applicable ethics regulation. Image reconstruction was performed using a recent 3D extension⁹ of algebraic image reconstruction based on an expanded signal model.^{26,27} In reconstructed images, residual weighting by net array sensitivity was removed by bias field correction (SPM12, <http://www.fil.ion.ucl.ac.uk/spm/software/spm12/>).²⁸ Background phase was removed using the STI Suite toolbox (<https://people.eecs.berkeley.edu/~chunlei.liu/software.html>), based on the V-SHARP algorithm.²⁹

3 | RESULTS

Figure 2 shows the results of the study described in Section 2.2, plotting total scan time against the underlying T_{AQ} per shot. First, this diagram illustrates the expected speed benefit of hexagonal sampling over conventional stacks in all resolution regimes. More importantly, it also confirms the benefit of tilting the hexagonal grids, which enhances the flexibility of how much k-space volume to cover per shot.

Supporting Information Figure S2 shows selected transverse slices from the first mono-planar T-Hex spiral-out scan, resolving 720-nanoliter voxels in 4.8 s. Data obtained with the same trajectories but stronger T_2^* weighting and sensitivity to susceptibility differences (TE = 25 ms) are shown




FIGURE 2 For three in-plane resolutions, the acquisition time for different stacks of spirals was computed and plotted against their total scan time. The latter comprises an overhead of 13 ms per shot, accounting for typical slab excitation, fat suppression and spoiling modules. For the trajectory design, a FOV of $24 \times 24 \times 12 \text{ cm}^3$ and $R = 8$ were assumed

in Figure 3, including image phase and enlarged details. Figure 4 presents slices from the mono-planar T-Hex EPI scan, achieving somewhat lower resolution, at 1 microliter per voxel, in only 4.3 s.

4 | DISCUSSION

Planar readouts along tilted hexagonal grids permit dividing the volume of a nominal k-space plane into fractions per shot other than inverse integers, while maintaining uniform sampling density. The key benefit of this approach is greater flexibility of optimizing the duration of each readout in 3D imaging. Such flexibility is instrumental not only for minimizing overall scan duration for given signal lifetime but also for optimizing contrast and SNR efficiency. Facilitating the use of multiple shots per nominal plane, the proposed strategy applies to the case of 3D imaging with relatively high resolution. The gain in flexibility is greatest in the regime of few shots per nominal plane where integer steps in the number of shots are the largest in relative terms.


Mono-planar T-Hex shares key properties with the kindred multi-planar T-Hex scheme, which enables fractional coverage of more than one nominal plane per shot for the case of moderate spatial resolution.⁹ Like the latter, mono-planar T-Hex spirals and EPI reconcile flexibility of segmentation with high average k-space speed, uniform density,

and smooth T_2^* weighting for time-efficient acquisition with near-optimal conditioning of reconstruction and tolerance to signal decay over long readouts. Unlike their multi-planar counterparts, the mono-planar readouts do not require blips to switch planes and potential slow-down to reduce associated sampling gaps.⁹

Demonstrated here with regular spiral-out and EPI trajectories, the proposed strategy is equally applicable to variations of these readouts such as inward spirals³⁰ and variable-density spirals³¹ and EPI.³² It is also expected to apply to multi-band imaging whose encoding requirements are essentially the same as for 3D imaging.³³

As illustrated in Figure 1D, the T-Hex sampling pattern repeats in the direction of the stack so that the sequence of readout trajectories is equally repetitive except for the prephaser gradient. As with regular stacks, this property may facilitate trajectory calibration, either by reducing the number of pre-measurements required or, in the case of concurrent trajectory recording with NMR probes, by reducing the rate of probe re-excitation in favor of T_1 recovery.²⁷ As also shown in Figure 1D, all tilting scenarios lead to hexagonal grids that each fall onto some finer Cartesian grid, which is not tilted with respect to the stack's principal axis. With EPI readouts, this property could be of utility for image reconstruction, permitting direct use of FFT without gridding, albeit with a FOV dictated by the embedding Cartesian grid. When this FOV is too small, it could be increased by refining the k-space grid by integer factors.

FIGURE 3 Slices from in-vivo data acquired with mono-planar T-Hex spiral-out. The whole brain is covered with $0.6 \times 0.6 \times 2 \text{ mm}^3$ resolution in 5.7 s, TE = 25 ms. The upper part shows magnitude images of six selected slices on the left, with corresponding phase images shown on the right. In the lower part, enlarged details are displayed



When it is too large, the excess size could offset the speed benefit of mere FFT. Alternatively, the primary hexagonal grid could be exploited by hexagonal fast Fourier transform (FFT) for image reconstruction, leading to hexagonal image representation.^{11,34} Either approach will require that the trajectories are played out highly accurately.

Modifications of fast 3D acquisition by relative rotation of spiral arms or shifting of echo-planar readouts have been explored in a number of previous contributions. Refs. [35,36] describe the heuristic selection of subsets of points on a regular Cartesian grid for favorable packing of neighboring interleaves, achieving approximate uniformity for specific degrees of segmentation. Approximate uniformity can also be achieved with spiral readouts with linear, heuristically chosen, or golden-angle increments of rotation as k_z increases.^{37,38} The T-Hex approach differs from these strategies in that it defines grids of uniform density for

variable segmentation. This relationship is connatural to that between schemes for covering multiple k -space planes per shot described in Refs. [39-41] and the multi-planar T-Hex approach.⁹ Approximately hexagonal or otherwise tailored patterns have also been explored for sampling domains other than common k -space such as k - t space for dynamic imaging,⁴² multi-contrast imaging,⁴³ or T_2 shuffling,⁴⁴ as well as k - k_y - t space for velocity spectroscopy⁴⁵ and k - k_{PSF} space for off-resonance correction.⁴⁶

For time efficiency, the proposed approach seeks to combine high average k -space speed with uniform sampling density for maximal acceleration within g -factor limits. Another strategy of containing g -factors is to start from strongly undersampling baseline trajectories and include oscillatory lateral excursions to reduce sampling gaps as, for example, in bunched phase encoding,⁴⁷ zig-zag sampling,⁴⁸ EPI or spirals with sinusoidal perturbation,^{49,50} and wave-CAIPI.^{51,52}




FIGURE 4 4 Mono-planar T-Hex EPI with $0.7 \times 0.7 \times 2 \text{ mm}^3$ nominal resolution and $\text{TE} = 20 \text{ ms}$, acquired in 4.3 s

However, such oscillation is limited by the feasible speed of gradient switching. For maximal gradient strength, G_{max} , and slew rate, S_{max} , the smallest radius of curvature that a trajectory can have at full speed is $\gamma G_{\text{max}}^2 / S_{\text{max}}$, which is typically in the same order of magnitude as the entire k-space range to be sampled. As a consequence, for oscillatory perturbations to accomplish substantial additional encoding, the underlying scan must exhibit relatively low k-space speed in the first place,⁵³ at the expense of overall scan time. For instance, whole-brain wave-CAIPI acquisition with resolution of $1 \times 1 \times 2 \text{ mm}^3$ and 9-fold acceleration has been reported to take 40 s,⁵¹ compared to 4.3 s with T-Hex EPI as reported here, achieving higher resolution ($0.7 \times 0.7 \times 2 \text{ mm}^3$) in the same FOV at eight-fold acceleration, with readouts of 33 ms vs. 14 ms. Orthogonal gradient oscillation can add a small amount of encoding capability also to amplitude-limited EPI and spirals, reflecting slight contributions to net k-space speed. However, such gain is ambivalent in that it relies on the fact that physically separate gradient coils and amplifiers are subject to individual current and voltage limitations. With angulation, orthogonal oscillating gradients will not yield the same benefit.

The benefits of mono-planar T-Hex acquisition in terms of scan speed, SNR yield, and contrast are greatest for 3D imaging at relatively high resolution. The bare speed advantage makes it particularly suitable for temporally resolved imaging like BOLD fMRI, ASL fMRI, and fQSM. With an emphasis rather on spatial resolution, it holds promise for fast structural gradient echo scans with plain T_2^* weighting, for SWI, and for QSM. The speed benefit is greatest for sequences with significant overhead for signal preparation such as MPRAGE and ASL, which readily

lend themselves to 3D EPI and spiral readouts. In addition, more efficient 3D encoding is attractive also for spin-echo imaging. Based on EPI and spiral readouts, GRASE (gradient and spin echo)⁵⁴⁻⁵⁶ is a natural area of application, with utility particularly for FLAIR (fluid-attenuated inversion recovery)⁵⁷ and ASL.^{58,59} However, mono-planar T-Hex should also lend itself to plain Cartesian 3D RARE (rapid acquisition with relaxation) imaging,^{60,61} exploiting even time progression in k-space for smooth T_2 weighting. Another spin-echo case of particular interest is diffusion-weighted imaging, which could benefit significantly due to its demand for spatial resolution and substantial overhead for diffusion weighting.⁶² Insofar as inter-shot phase inconsistency due to microscopic motion during diffusion weighting can be managed, for example, by navigation,^{63,64} array acquisition⁶⁵ or compensation,⁶⁶ diffusion imaging is amenable to 3D acquisition and T-Hex sampling. For all potential applications, high B_0 strengthens the case by boosting baseline sensitivity in favor of targeting high resolution at high speed.

ACKNOWLEDGMENTS

We thank Roger Lüchinger for technical support, Markus Weiger for valuable insights regarding SNR considerations, and Emily Baadsvik for her linguistic expertise. Technical support from Philips Healthcare is gratefully acknowledged.

CONFLICT OF INTEREST

Klaas Pruessmann holds a research agreement with and receives research support from Philips. He is a shareholder of Gyrotools LLC. Bertram Wilm and Benjamin Dietrich are employees of Skope Magnetic Resonance Technologies AG.

ORCID

Maria Engel  <https://orcid.org/0000-0002-0143-1908>

Lars Kasper  <https://orcid.org/0000-0001-7667-603X>

Bertram Wilm  <https://orcid.org/0000-0002-8577-8182>

Franz Patzig  <https://orcid.org/0000-0002-2418-5893>

Klaas P. Pruessmann  <https://orcid.org/0000-0003-0009-8362>

REFERENCES

- Feinberg DA, Moeller S, Smith SM, et al. Multiplexed echo planar imaging for sub-second whole brain fMRI and fast diffusion imaging Valdes-Sosa PA, editor. *PLoS One*. 2010;5:e15710.
- Borogovac A, Asllani I. Arterial spin labeling (ASL) fMRI: advantages, theoretical constraints and experimental challenges in neurosciences. *Int J Biomed Imaging*. 2012;2012:1-13. <https://doi.org/10.1155/2012/818456>
- Vidorreta M, Wang Z, Rodríguez I, Pastor MA, Detre JA, Fernández-Seara MA. Comparison of 2D and 3D single-shot ASL perfusion fMRI sequences. *Neuroimage*. 2013;662-671.
- Balla DZ, Sanchez-Panchuelo RM, Wharton SJ, et al. Functional quantitative susceptibility mapping (fQSM). *Neuroimage*. 2014;100:112-124.
- Langkammer C, Bredies K, Poser BA, et al. Fast quantitative susceptibility mapping using 3D EPI and total generalized variation. *Neuroimage*. 2015;111:622-630.
- Fernández-Seara MA, Wang ZE, Wang J, et al. Continuous arterial spin labeling perfusion measurements using single shot 3D GRASE at 3 T. *Magn Reson Med*. 2005;54:1241-1247.
- Poser BA, Koopmans PJ, Witzel T, Wald LL, Barth M. Three dimensional echo-planar imaging at 7 Tesla. *Neuroimage*. 2010;51:261-266.
- Yang Y, Glover GH, van Gelderen P, et al. Fast 3D functional magnetic resonance imaging at 1.5 T with spiral acquisition. *Magn Reson Med*. 1996;36:620-626.
- Engel M, Kasper L, Wilm B, et al. T-Hex: tilted hexagonal grids for rapid 3D imaging. *Magn Reson Med*. 2021;85:2507-2523.
- Thue A. Über die dichteste Zusammenstellung von kongruenten Kreisen in einer Ebene. *Christiania Vid. Selsk. Skr*. 1910;1:3-9.
- Mersereau RM. The processing of hexagonally sampled two-dimensional signals. *Proc IEEE*. 1979;67:930-949.
- Ehrhardt JC. MR data acquisition and reconstruction using efficient sampling schemes. *IEEE Trans Med Imaging*. 1990;9:305-309.
- Kim WS, Cho Z-H. The application of hexagonal sampling in NMR imaging. In: Proceedings of the International Society for Magnetic Resonance in Medicine. New York, NY; 1990. p. 551.
- Jurrissen M, Fuderer M & van den Brink J Diamond-SENSE: undersampling on a crystallographic grid. In: Proceedings of the 12th Annual Meeting of ISMRM. Vol. 2643. Kyoto, Japan; 2004.
- Breuer FA, Blaimer M, Mueller MF, et al. Controlled aliasing in volumetric parallel imaging (2D CAIPIRINHA). *Magn Reson Med*. 2006;55:549-556.
- Saranathan M, Ramanan V, Gulati R, Venkatesan R. ANTHEM: anatomically tailored hexagonal MRI. *Magn Reson Imaging*. 2007;25:1039-1047.
- Sveinsson B, Worters PW, Gold GE, Hargreaves BA. Hexagonal undersampling for faster MRI near metallic implants. *Magn Reson Med*. 2015;73:662-668.
- Hennel F. Multiple-shot echo-planar imaging. *Concept Magn Reson*. 1997;9:43-58.
- Lustig M, Kim S-J, Pauly JM. A fast method for designing time-optimal gradient waveforms for arbitrary k-space trajectories. *IEEE Trans Med Imaging*. 2008;27:866-873.
- Kaldoudi E, Williams SCR, Barker GJ, Tofts PS. A chemical shift selective inversion recovery sequence for fat-suppressed MRI: theory and experimental validation. *Magn Reson Imaging*. 1993;11:341-355.
- Mason GF, Harshbarger T, Hetherington HP, Zhang Y, Pohost GM, Twieg DB. A Method to measure arbitrary k-space trajectories for rapid MR imaging. *Magn Reson Med*. 1997;38:492-496.
- Börnert P, Schomberg H, Aldefeld B, Groen J. Improvements in spiral MR imaging. *Magn Reson Mater Phys Biol Med*. 1999;9:29-41.
- De Zanche N, Barmet C, Nordmeyer-Massner JA, Pruessmann KP. NMR probes for measuring magnetic fields and field dynamics in MR systems. *Magn Reson Med*. 2008;60:176-186.
- Gross S, Barmet C, Dietrich BE, Brunner DO, Schmid T, Pruessmann KP. Dynamic nuclear magnetic resonance field sensing with part-per-trillion resolution. *Nat Commun*. 2016;7:13702.
- Dietrich BE, Brunner DO, Wilm BJ, et al. A field camera for MR sequence monitoring and system analysis. *Magn Reson Med*. 2016;75:1831-1840.
- Kasper L, Haeberlin M, Dietrich BE, et al. Matched-filter acquisition for BOLD fMRI. *Neuroimage*. 2014;100:145-160.
- Kasper L, Engel M, Barmet C, et al. Rapid anatomical brain imaging using spiral acquisition and an expanded signal model. *Neuroimage*. 2017;168:88-100.
- Ashburner J, Friston KJ. Unified segmentation. *Neuroimage*. 2005;26:839-851.
- Li W, Wu B, Liu C. Quantitative susceptibility mapping of human brain reflects spatial variation in tissue composition. *Neuroimage*. 2011;55:1645-1656.
- Peter B, Bernd A, Holger E. Reversed spiral MR imaging. *Magn Reson Med*. 2000;44:479-484. [https://doi.org/10.1002/1522-2594\(200009\)44:3<479::aid-mrm20>3.0.co;2-2](https://doi.org/10.1002/1522-2594(200009)44:3<479::aid-mrm20>3.0.co;2-2)
- Spielman DM, Pauly JM, Meyer CH. Magnetic resonance fluoroscopy using spirals with variable sampling densities. *Magn Reson Med*. 1995;34:388-394.
- Zeller M, Müller A, Gutberlet M, et al. Boosting BOLD fMRI by K-space density weighted echo planar imaging. *PLoS One*. 2013;8:e74501.
- Zahneisen B, Poser BA, Ernst T, Stenger VA. Three-dimensional Fourier encoding of simultaneously excited slices: generalized acquisition and reconstruction framework. *Magn Reson Med*. 2014;71:2071-2081.
- Rummelt NI, Wilson JN. Array set addressing: enabling technology for the efficient processing of hexagonally sampled imagery. *J Electron Imaging*. 2011;20:023012. <https://doi.org/10.1117/1.3589306>
- Deng W, Zahneisen B, Stenger VA. Rotated stack-of-spirals partial acquisition for rapid volumetric parallel MRI. *Magn Reson Med*. 2016;76:127-135.
- Hendriks AD, D'Agata F, Raimondo L, et al. Pushing functional MRI spatial and temporal resolution further: high-density receive arrays combined with shot-selective 2D CAIPIRINHA for 3D echo-planar imaging at 7 T. *NMR Biomed*. 2020;33:e4281.
- Anderson AG, Wang D, Pipe JG. Controlled aliasing for improved parallel imaging with a 3D spiral staircase trajectory. *Magn Reson Med*. 2020;84:866-872.

38. Turley DC, Pipe JG. Distributed spirals: a new class of three-dimensional k-space trajectories. *Magn Reson Med.* 2013;70:413-419.
39. Narsude M, van der Zwaag W, Kober T, Gruetter R, Marques JP. Improved temporal resolution for functional studies with reduced number of segments with three-dimensional echo planar imaging. *Magn Reson Med.* 2014;72:786-792.
40. Narsude M, Gallichan D, van der Zwaag W, Gruetter R, Marques JP. Three-dimensional echo planar imaging with controlled aliasing: a sequence for high temporal resolution functional MRI. *Magn Reson Med.* 2016;75:2350-2361.
41. Stirnberg R, Stöcker T. Segmented K-space blipped-controlled aliasing in parallel imaging for high spatiotemporal resolution EPI. *Magn Reson Med.* 2021;85:1540-1551. <https://doi.org/10.1002/mrm.28486>
42. Tsao J, Kozerke S, Boesiger P, Pruessmann KP. Optimizing spatiotemporal sampling for k-t BLAST and k-t SENSE: application to high-resolution real-time cardiac steady-state free precession. *Magn Reson Med.* 2005;53:1372-1382.
43. Wang F, Dong Z, Reese TG, et al. Echo planar time-resolved imaging (EPTI). *Magn Reson Med.* 2019;81:3599-3615.
44. Tamir JJ, Uecker M, Chen W, et al. T2 shuffling: sharp, multi-contrast, volumetric fast spin-echo imaging. *Magn Reson Med.* 2017;77:180-195.
45. Hansen MS, Baltus C, Tsao J, et al. Accelerated dynamic Fourier velocity encoding by exploiting velocity-spatio-temporal correlations. *MAGMA.* 2004;17:86-94.
46. Dong Z, Wang F, Reese TG, et al. Tilted-CAIPI for highly accelerated distortion-free EPI with point spread function (PSF) encoding. *Magn Reson Med.* 2019;81:377-392.
47. Moriguchi H, Sunshine JL & Duerk JL Further scan time reduction of bunched phase encoding using sensitivity encoding. In: Proceedings of the 13th Annual Meeting of ISMRM. Miami, Florida, USA; 2005. p. 287.
48. Breuer FA, Moriguchi H, Seiberlich N, et al. Zigzag sampling for improved parallel imaging. *Magn Reson Med.* 2008;60:474-478.
49. Haeberlin M, Wilm B, Barmet C, Kozerke S, Katsikatsos G, Pruessmann KP. Sinusoidal perturbations improve the noise behavior in parallel EPI. In: Proceedings of the 18th Annual Meeting of ISMRM. Stockholm, Sweden; 2010. p. 2872.
50. Chu A, Noll DC. Simultaneous multislice spiral imaging using z-gradient modulation and parallel receive coils. In: Proceedings of the 21st Annual Meeting of ISMRM. Salt Lake City, UT, USA; 2013. p. 3317.
51. Bilgic B, Gagoski BA, Cauley SF, et al. Wave-CAIPI for highly accelerated 3D imaging. *Magn Reson Med.* 2015;73:2152-2162.
52. Poser BA, Bilgic B, Gagoski BA, et al. Echo-planar imaging with wave-CAIPI acquisition and reconstruction. In: Proceedings of the 25th Annual Meeting of ISMRM. Honolulu, HI, USA; 2017. p. 1198.
53. Poser BA, Setsompop K. Pulse sequences and parallel imaging for high spatiotemporal resolution MRI at ultra-high field. *Neuroimage.* 2018;168:101-118.
54. Oshio K, Feinberg DA. GRASE (Gradient-and Spin-Echo) imaging: a novel fast MRI technique. *Magn Reson Med.* 1991;20:344-349.
55. Mugler JP. Optimized three-dimensional fast-spin-echo MRI. *J Magn Reson Imaging.* 2014;39:745-767.
56. Fielden SW, Mugler JP, Hagspiel KD, Norton PT, Kramer CM, Meyer CH. Noncontrast peripheral MRA with spiral echo train imaging. *Magn Reson Med.* 2015;73:1026-1033.
57. Li Z, Pipe JG, Ooi MB, Kuwabara M, Karis JP. Improving the image quality of 3D FLAIR with a spiral MRI technique. *Magn Reson Med.* 2020;83:170-177.
58. Vidorreta M, Balteau E, Wang Z, et al. Evaluation of segmented 3D acquisition schemes for whole-brain high-resolution arterial spin labeling at 3 T. *NMR Biomed.* 2014;27:1387-1396. <https://doi.org/10.1002/nbm.3201>
59. Dai W, Garcia D, de Bazelaire C, Alsop DC. Continuous flow-driven inversion for arterial spin labeling using pulsed radio frequency and gradient fields. *Magn Reson Med.* 2008;60:1488-1497.
60. Hennig J, Nauert A, Friedburg H. RARE imaging: a fast imaging method for clinical MR. *Magn Reson Med.* 1986;3:823-833.
61. Yuan C, Schmiedl UP, Weinberger E, Krueck WR, Rand SD. Three-dimensional fast spin-echo imaging: pulse sequence and in vivo image evaluation. *J Magn Reson Imaging.* 1993;3:894-899.
62. Frank LR, Jung Y, Inati S, Tyszka JM, Wong EC. High efficiency, low distortion 3D diffusion tensor imaging with variable density spiral fast spin echoes (3D DW VDS RARE). *Neuroimage.* 2010;49:1510-1523.
63. Liu C, Bammer R, Kim D, Moseley ME. Self-navigated interleaved spiral (SNAILS): application to high-resolution diffusion tensor imaging. *Magn Reson Med.* 2004;52:1388-1396.
64. Steinhoff M, Nehrke K, Mertins A, Börner P. Segmented diffusion imaging with iterative motion-corrected reconstruction (SEDIMENT) for brain echo-planar imaging. *NMR Biomed.* 2020; 33:e4185.
65. Chen N, Guidon A, Chang H-C, Song AW. A robust multi-shot scan strategy for high-resolution diffusion weighted MRI enabled by multiplexed sensitivity-encoding (MUSE). *Neuroimage.* 2013;72:41-47.
66. Stoek CT, von Deuster C, Genet M, Atkinson D, Kozerke S. Second-order motion-compensated spin echo diffusion tensor imaging of the human heart. *Magn Reson Med.* 2016;75:1669-1676.

SUPPORTING INFORMATION

Additional supporting information may be found online in the Supporting Information section.

FIGURE S1 SNR in dependence of T_2^* and acquisition time per shot. The orange line marks those acquisition times which maximize the SNR for a given T_2^*

FIGURE S2 Images acquired with T-Hex spiral-out. The whole brain is covered with $0.6 \times 0.6 \times 2 \text{ mm}^3$ resolution in 4.8 s, TE = 15 ms. Red lines mark the position of the displayed slices

How to cite this article: Engel M, Kasper L, Wilm B, et al. Mono-planar T-Hex: Speed and flexibility for high-resolution 3D imaging. *Magn Reson Med.* 2022;87:272–280. <https://doi.org/10.1002/mrm.28979>